

CORRIGENDUM

Tender Ref. No. IITK/NAF/TENDER/2022-23/21

February 23, 2023

1. Page 3, Table 1

Existing

Sl. No.	Item Name	Qty	Mounting	Dimensions	Weight	Power Requirements
3	HR-ToF-AMS without Desktop computer and without rack	1	Rack	41inch (width) x 24inch (depth) x 53inch (height)	150Kg	220VAC, 50Hz, 600W
4	Desktop computer for HR-ToF-AMS	1	Rack	Standard cabinet + 15.4inch monitor	-	220VAC, 50Hz

Revised

Sl. No.	Item Name	Qty	Mounting	Dimensions	Weight	Power Requirements
3& 4	HR-ToF-AMS with Desktop computer and rack	1	Rack*	41inch (width) x 24inch (depth) x 53inch (height) and Standard cabinet with 15.4inch monitor on top of stand	150Kg	220VAC, 50Hz, 600W (for equipment) and 220VAC, 50Hz (for desktop computer)

*The rack provided by IIT-Kanpur needs to be fixed by supplier inside the mobile air quality monitoring laboratory during delivery.

2. Page 5, Table 2

Sl.no.4, 'Rack-2 for HR-ToF-AMS with computer' stands removed from Table-2. The rack will be provided by IIT-Kanpur as mentioned above.

3. Page 5, Table 2

The following points are added in Table 2

Sl. No.	Item Name	Quantity
29	Platform for PM+Gas sensor {size: 31cm (length) x 17cm (width) x 2.5cm (height)}**	2
30	Provision for strapping of items on roof-top	NA

** Each platform should not have any obstruction along one side of the width.

Sachchida Naw Tripathi

Sachchida Tripathi