

## CORRIGENDUM-I

Date: 11-09-2024

**Enquiry No: IITK/NAF/TENDER/2024-25/05**

**Continuous Automatic Particulate Monitor**

Please note the below-listed changes in Technical Specifications.

	<b>Existing</b>	<b>To be read as</b>
Measurement Object	PM10 or PM2.5 (preferred simultaneous measurements)	Both PM2.5 and PM10 sampling required. Simultaneous sampling preferred. Instrument able to sample either of PM2.5 and PM10 will also be considered.

The following is added to technical specifications.

Housing / Mounting	Should be mountable on rack / tripod / table-top. Bidder should mention the type of housing/mounting in Technical Bid and quote as separate line item in Financial Bid.
--------------------	--

*Sachchida Nand Tripathi*

Sachchida Nand Tripathi

**S N TRIPATHI**  
**Coordinator**  
**National Aerosol Facility**  
**IIT Kanpur - 208016, INDIA**