

**Indian Institute of Technology-Kanpur**  
**Department of Biological Sciences & Bioengineering**

Enquiry No.: IITK/BSBE/JGR- 5620/2015-16/06

Enquiry date: 10-06-15

Last Date: 26-06-15

**Enquiry for the Refrigeration bath circulator**

We would like to purchase a Refrigeration bath circulator for our laboratory. Offered product should take into account specifications given below.

**Specification**

- 1- 6-10 L BATH VOLUME
- 2- Temp Range: -28 to +150 DEGREE C,
- 3- Temperature stability (°C): 0.01
- 4- Cooling Capacity at +20°C: 320 Watts
- 5- Heating Capacity: 2kW
- 6- Volume liters: 6 -10
- 7- Power Requirements: 230V 50Hz
- 8- Pump Flow Rate: 20lpm@475mbar
- 9- Maximum Suction: 330mbar

**Terms and Conditions:**

- Cost of the items including relevant accessories with technical specification should be mentioned in detail.
- Prices should include delivery up to IIT Kanpur.
- Validity of quotation should be at least for 90 days.
- Delivery period: within 30days from the date of purchase order kindly send your best offer (Technical offer separately) and the quotation along with your authorized dealer certificate to address given below.
- Comprehensive warranty for 3 years should be supplied.

  
Dr. Jayandharan G. Rao

Dr. Jayandharan G. Rao, Ph.D.  
Associate Professor  
Department of Biological  
Sciences and Bioengineering  
भारतीय प्रौद्योगिकी संस्थान, कानपुर  
Indian Institute of Technology, Kanpur  
Kanpur-208016, U.P. (INDIA)