

Accelerometer Calibration Certificate

Order Code : 060-J077-04
Type : 7100-10
Serial No : 351955
Sensitivity : 4.48 mV/g
Bias Voltage : -0.02 mV
Mounting : 4 Point
Cable Length : 5m 4 Core Screened O/B PVC

Connection Details

Red : +5V
Yellow : Signal Hi
Blue : 0 Volts
Green : Signal Lo
Screen : Not Connected To Case

Works Order No: 41022

Date: 26/02/2009

Tested By: S.Collins

Signature: *S. A. Collins*

Certified that the instrument has been calibrated using equipment certified and traceable to UK national standards carried out in accordance with the general requirements of BS EN ISO 10012-1. The instrument complies with the manufacturer's specifications.