

Accelerometer Calibration Certificate

Order Code : 060-J077-04
Type : 7100-10
Serial No : 351960
Sensitivity : 4.40 mV/g
Bias Voltage : 0.01 mV
Mounting : 4 Point
Cable Length : 5m 4 Core Screened O/B PVC

Connection Details

Red : +5V
Yellow : Signal Hi
Blue : 0 Volts
Green : Signal Lo
Screen : Not Connected To Case

Works Order No: 41022 Date: 26/02/2009

Tested By: S.Collins

Signature: _____

S. A. Collins

Certified that the instrument has been calibrated using equipment certified and traceable to UK national standards carried out in accordance with the general requirements of BS EN ISO 10012-1. The instrument complies with the manufacturer's specifications.