

Accelerometer Calibration Certificate

Type: MTN/7100-10
Serial No.: 333468
Range: +/- 10 g
Sensitivity: 2.84 mV/g
Offset Bias: 0.03 mV
Mounting: Small 4 Point
Cable: 5M O/B Cable PVC

Connection Details

Red: +ve Supply (5 Volts Only)
Blue: 0 Volts
Yellow: +ve Output
Green: -ve Output

Works Order No: 38612 Date: 11th February 2008

Tested by: S.Collins Signature: *S. Collins*

**Certified that the instrument has been calibrated using equipment certified and traceable to UK National Standards carried out in accordance with the general requirements of BS EN ISO 10012-1
The instrument complies with the manufacturer's specifications**