

Accelerometer Calibration Certificate

Type: MTN/7100-10
Serial No.: 333469
Range: +/- 10 g
Sensitivity: 2.68 mV/g
Offset Bias: 0.02 mV
Mounting: Small 4 Point
Cable: 5M O/B Cable PVC

Connection Details

Red: +ve Supply (5 Volts Only)
Blue: 0 Volts
Yellow: +ve Output
Green: -ve Output

Works Order No: 38612
Tested by: S.Collins

Date: 11th February 2008
Signature: *S. Collins*

**Certified that the instrument has been calibrated using equipment certified and traceable to UK National Standards carried out in accordance with the general requirements of BS EN ISO 10012-1
The instrument complies with the manufacturer's specifications**