

# Accelerometer Calibration Certificate

Type: MTN/7100-10  
Serial No.: 333471  
Range: +/- 10 g  
Sensitivity: 2.56 mV/g  
Offset Bias: 0.03 mV  
Mounting: Small 4 Point  
Cable: 5M O/B Cable PVC

## Connection Details

Red: +ve Supply (5 Volts Only)  
Blue: 0 Volts  
Yellow: +ve Output  
Green: -ve Output

Works Order No: 38612  
Tested by: S.Collins

Date: 11th February 2008  
Signature: *S. Collins*

**Certified that the instrument has been calibrated using equipment certified and traceable to UK National Standards carried out in accordance with the general requirements of BS EN ISO 10012-1**  
**The instrument complies with the manufacturer's specifications**