

# Accelerometer Calibration Certificate

Type:	MTN/7100-10
Serial No.:	333472
Range:	+/- 10 g
Sensitivity:	2.92 mV/g
Offset Bias:	0.05 mV
Mounting:	Small 4 Point
Cable:	5M O/B Cable PVC

## Connection Details

Red:	+ve Supply (5 Volts Only)
Blue:	0 Volts
Yellow:	+ve Output
Green:	-ve Output

Works Order No:	38612	Date:	11th February 2008
Tested by:	S.Collins	Signature:	<i>S. Collins</i>

**Certified that the instrument has been calibrated using equipment certified and traceable to UK National Standards carried out in accordance with the general requirements of BS EN ISO 10012-1  
The instrument complies with the manufacturer's specifications**