

# Honeywell

**Honeywell**  
 Sensotec Sensors  
 2080 Arlingate Lane  
 Columbus, Ohio 43228 U.S.A.  
 Phone: 614-850-5000  
 Fax: 614-850-1111  
 URL: <http://www.sensotec.com>  
 E-mail: [service@sensotec.com](mailto:service@sensotec.com)

## Certificate of Calibration

### Product Identification

Product Type: Load Cell  
 Serial Number: 1271803

Model: 43  
 Part Number: 060-2716-10  
 Order Code: AL112MX

### Product Specifications

Full Scale Range: 1500 kN  
 Calibrated At: 1500 kN  
 Compression

Excitation: 10. VDC  
 Input Impedance: 371  $\Omega$   
 Output Impedance: 243  $\Omega$   
 Electrical Leakage:  $\infty$  Meg $\Omega$

### Calibration Data

Calibration Factor: 2.9984 mV/V

Shunt Cal Factor: 1.0246 mV/V  
 Shunt Cal Resistor: 59k $\Omega$

### Wiring Code

UNAMP#2,4-COND,6-PIN	
PIN	DESIGNATION
A	(+)EXCITATION
B	(+)EXCITATION
C	(-)EXCITATION
D	(-)EXCITATION
E	(-)OUTPUT
F	(+)OUTPUT

001-0333-02

### Certification Information

Type of Calibration: Standard  
 Calibration Date: 04/03/2009

Certificate Number: 086-0000-00  
 Calibration Procedure: 072-LC75-10

*Notes: Instruments used in the calibration of this product have been calibrated to standards traceable to the National Institute of Standards and Technology (NIST). A calibration uncertainty ratio of 4:1 has been maintained unless otherwise stated.*

*If you have any questions concerning this certificate of calibration, or for recalibration or repair of this product, please visit our website at <http://www.sensotec.com/service.htm> or call our service department at (614) 850-5000.*

*This is a quality record.*