

# SENSOTEC

2080 ARLINGATE LANE COLUMBUS, OHIO 43228 (614) 850 - 5000  
INTERNET URL <http://www.sensotec.com>

## CERTIFICATE OF CALIBRATION

MODEL: TH/1590-05  
ORDER CODE: BL911MX  
SERIAL NUMBER: 967552  
CALIBRATION DATE: Aug 21/2003  
INPUT RESISTANCE: 352.Ω  
OUTPUT RESISTANCE: 352.Ω  
LEAKAGE: ∞


CAPACITY: 500 KN  
COMPRESSION  
CALIBRATED AT: 500 KN  
EXCITATION: 10.0 VDC  
CALIBRATION FACTOR: 2.008 MV/V  
SHUNT RESISTOR: 59kΩ  
SHUNT CAL FACTOR: 1.487 MV/V

WIRING CODE

UNAMP#1,4-COND,CBL

COLOR	DESIGNATION
RED	(+)EXCITATION
BLACK	(-)EXCITATION
GREEN	(-)OUTPUT
WHITE	(+)OUTPUT

001-0333-01



Accepted and Certified by: Michael A Stanley  
Date Printed: 8/22/2003