

553

SENSOTEC

UNAME
COLOR TE LANE, OHIO 43228 (614) 850 - 5000
RED http://www

BLACK
GREEN
WHITE

CIFICATE OF CALIBRATION

MTH/1590-05
ORDER BL911MX
AL NU 967553
TION Aug 21/2003
RESIST 352.Ω
RESIST 352.Ω
LEA ∞

CAPACITY: 500 KN
COMPRESSION
CALIBRATED AT: 500 KN
EXCITATION: 10.0 VDC
CALIBRATION FACTOR: 2.017 MV/V
SHUNT RESISTOR: 59kΩ
SHUNT CAL FACTOR: 1.487 MV/V




380 ARLINGA
INTERNET URL

P#1,4-C
DES
(+)EIN
(-)EIN
(-)OI
(+)OI

SERIAL
CALIBRA
INPUT R
OUTPUT R

001-0333-01



Accepted and Certified by: Michael A Stanley
Date Printed: 8/22/2003