

# SENSOTEC

2080 ARLINGATE LANE COLUMBUS, OHIO 43228 (614) 850-5000  
INTERNET URL <http://www.sensotec.com>

## CERTIFICATE OF CALIBRATION

MODEL:	TH/1590-05	CAPACITY:	500 KN
ORDER CODE:	BL911MX		COMPRESSION
SERIAL NUMBER:	967554	CALIBRATED AT:	500 KN
CALIBRATION DATE:	Aug 21/2003	EXCITATION:	10.0 VDC
INPUT RESISTANCE:	352.Ω	CALIBRATION FACTOR:	2.013 MV/V
OUTPUT RESISTANCE:	352.Ω	SHUNT RESISTOR:	59kΩ
LEAKAGE:	∞	SHUNT CAL FACTOR:	1.487 MV/V
WIRING CODE			

UNAMP#1,4-COND,CBL

COLOR	DESIGNATION
RED	(+)EXCITATION
BLACK	(-)EXCITATION
GREEN	(-)OUTPUT
WHITE	(+)OUTPUT

001-0333-01



SENSOTEC

Accepted and Certified by:

*Michael A Stanley*

Date Printed: 8/22/2003