

Honeywell


2080 ARLINGATE LANE COLUMBUS, OHIO 43228 (614) 850 - 5000

CERTIFICATE OF CALIBRATION

MODEL: JEC-AG
PART NO.: 060-H032-03
ORDER CODE: AY322HR
SERIAL NUMBER: L12451200
CALIBRATION DATE: 02/27/2009

RANGE: +/-2.000 INCHES
LINEARITY: 0.2500%
SUPPLY: 40.0 VDC
CALIBRATION FACTOR: 2.4780 V/INCH

WIRING CODE



Accepted and Certified by:

Eszter Gozon

Date Printed: 3/11/2009

Eszter Gozon, Quality Manager