

## Request for quotation

Enquiry number: AE/DPM/Massflowmeter/2015/3

Enquiry date: 4<sup>th</sup> Aug 2015

Closing Date: 16<sup>th</sup> August 2015

Sealed quotations should reach the undersigned latest by **(date)** for the following:

Serial #	Description	Quantity
1	Mass flow meter (air), 0-2500 LPM	1
2	Mass flow meter (CNG), 0-250 LPM	1

The above-mentioned equipment should conform to the following specifications:

- Operating Temperature : 10 to 50 °C
- Inlet pressure : 2 bar
- Accuracy : 0.2% or better

Any additional accessory necessary for operation of the above mass flow controllers, in two different applications simultaneously, such as power supplies, communication software, cables, etc. must be included and clearly specified as essential or optional.

### Terms and conditions:

- Quotations should have a validity of minimum of 60 days
- The equipment should be provided with a warranty of at least two years after installation
- The delivery period should be specifically stated.
- Permissible educational discount should be provided since the equipment will be used for research work of students.
- For suppliers from outside India, the rate offered should be FOB (specify city) or FCA terms.

### Address for the quotation:

Dr. DP Mishra  
Combustion Laboratory  
Department of Aerospace Engineering  
Indian Institute of Technology  
Kanpur-208016, India  
E-mail: mishra@iitk.ac.in  
Phone: +91-512-2597125