

Revised Enquiry No: ME /KKK/03R2/2017-18

Dated: June 07, 2017

Quotation (Technical and commercial , separately) must reach to us till **16.06.2017** before 5.00PM and should be sent to Prof. Kamal K. Kar, Department of Mechanical Engineering, Room # ACMS 203D, IIT-Kanpur, 208016 against the enquiry letter numbered **ME /KKK/03R2/2017-18 Dated: June, 07, 2017.**

Dear Sir/Madam:

Quotations (Only Technical **(in this format supported by printed leaflet and all other detailed technical specifications are to be quoted separately)** and commercial) are invited for purchase of "High temperature Hall Measurement System" having following specifications:

Specifications: 1(one) unit High temperature Hall Measurement System having Detailed General Features:

Type of magnet: electromagnet	Electromagnet
Specification of magnet (Tesla)	More than 1 Tesla, water cooling facility,
Pole diameter, pole gap	Please specify for your system
Power supply for magnet (voltage, frequency, etc)	All power supplies should be configured for 220V/50Hz.
Chiller, specification of chiller	2 L/min, 1 bar , minimum (should be arranged by IIT-Kanpur, no need to quote it)
Vacuum	7 mTorr or better (should be arranged by IIT-Kanpur, no need to quote it)
Environmental control chamber	7 mTorr or better
Temperature range	300K - 700K Please quote separately if you have 1000K or more than 700K
Specifications of sample holder if any	four probes for van der Pauw measurements or Kapton Harness for contacts soldering
Sample stage material	Ceramics/special glass
Any reference sample for calibration	Specify as per your system
Range of Resistance (or Resistivity): Ohm*cm	Specify for your system
Range of Carrier Concentration:cm ⁻³	Specify for your system
Range of Mobility :cm ² /volt*sec	Specify for your system
Temp uniformity	0.1 K or better (depends on wait time)
Any other materials parameters that can be extracted from this measurement	Specify if you have any
Specification for software	Windows 7, 8 or 10, English based operating System (Computer will be arranged by IIT Kanpur, software will be installed in the supplied computer only)
Upgradation of software	Specify your system
Printer	No need printer
Is the measurement step function at a fixed temp or a continuous function with temperature	Specify for your system
Plotting graph of Concentration vs temp	Specify for your system
Plotting graph of mobility vs temp	Specify for your system
Plotting graph of resistivity vs temp	Specify for your system
Plotting graph of hall coefficient vs temp	Specify for your system
Any other parameters, if any	Specify for your system
Special accessories	Specify for your system
Consumables	Specify for your system

Other Conditions:

1. Prices should mention FOB (your international airport), CIP New Delhi, and IIT-Kanpur including Packing and Forwarding, Insurance and freight.

2. **Warranty for three years without any extra charges**

3. Validity of quotation should be at least for 90 days.

4. Maximum educational discount

5. Any other charges, if applicable.

6. User list of this valve (please give complete address, including cell number, email address)

7. Proprietary certificate, if any

8. Authorization letter from your principal, if any

9. Agency commission, if any

10. Printed technical literatures as per your specifications (Xerox copy will not be accepted) and website for your product if any

11. Currency conversion that will be considered as 1US\$=70 INR, 1GBP=89 INR, 1EURO=78 INR

Kindly mention **ME /KKK/03R2/2017-18 Dated: June 07, 2017** on envelope carrying quotation and printed literature and send your offer (Technical and commercial, separately) so as to reach us on or before June 16, 2017 to the following address-

Please note: Your technical bid will be disqualified, if you don't quote as per suggested format and missing of any technical points

Prof. Kamal K. Kar,

Department of Mechanical Engineering

Room # 203 ACMS

IIT Kanpur - 208016 India

You will be informed accordingly by appropriate authorities if your bid is accepted. Please do not make unnecessary and unsolicited phone calls and emails (except to seek a clarification for the tender), you will be contacted as per need.