

Corrigendum for Enquiry No. IITK/BB/2018/Acce

Due to some typing mistake in previous tender enquiry, IITK/BB/2018/Acce of Seismic, Miniature IPC based Accelerometers the specifications should read as follows:

Specification	Model 01	Model 02
Measurement range	± 50 g pk (490 m/s ² pk)	0.5 g pk
Sensitivity ($\pm 10\%$)	100 mV/g or higher	10000 mV/g
Frequency range ($\pm 5\%$)	($\pm 5\%$) 1.0 to 8000 Hz	($\pm 5\%$) 0.15 to 1000 Hz
Resonant Frequency	≥ 30 kHz	≥ 10 kHz
Overload limit(Shock)	± 5000 g pk	± 5000 g pk
Constant Current Excitation	2 to 20 mA	2 to 20 mA
Size	5 mm x 13 mm x 8 mm (h x l x w)	32 mm x 56 mm (hex x height)
Mounting	Adhesive	1/4-28 Female Thread
Electrical Connection Position	Side	Top
Weight	≤ 1.0 gm	≤ 250 gm

Extension of the last date for tender submission: Vendors may have to alter their offer due to the amendment. Hence the tender submission date is being extended 1 week from earlier 09/04/2018. The revised last date for tender submission is 16/04/2018 and revised bid opening date is 17/04/2018.

Resubmission of tender document: The vendors who have already submitted their offer will be allowed to resubmit due to this amendment. The last date as mentioned above is also applicable to them. In case of resubmission of bid earlier bid will be considered invalid and will not be opened. The vendor will be evaluated based on the latest offer. If there is no resubmission from the vendor, then the vendor will be evaluated based on the submitted bid.



Prof. B. Bhattacharya
Dept. of Mechanical Engg.
Indian Institute of Technology Kanpur
Kanpur 208016
E-Mail: bishakh@iitk.ac.in
Phone: 05122597913