

Indian Institute of Technology - Kanpur
Department of Biological Sciences & Bioengineering

Enquiry Number: **JS/BSBE/2017-2018/01**, dated: **12.05.2017**

Sub.: Inquiry for the Refrigerated table top centrifuge

Opening date: 12 May, 2017 at 10:00 AM

Closing date: 24 May, 2017 at 5:00 PM

Sealed quotes (technical bid and price bid separately sealed) are invited for the **Refrigerated table top centrifuge**, as per the specifications given in the next page.

***Your quote should mention/include the following:**

- Maximum discount if any should be offered and mentioned.
- Quoted price should include the cost for installation, warranty and required accessories (see below).
- Validity of the quote at least for 90 days.
- FOB (indicating port of shipment) and CIF (New Delhi) values should be quoted separately if import is required. For quotes in INR, the price quote should be for delivery at Kanpur.
- The quote should cover insurance for transport up to Kanpur.
- Indian agency commission if applicable (should be certified by the principal if no agency commission is applicable) in case of import.
- Authorization certificate from the principal if you are a local agent.
- Terms and conditions for the payment, including the banker's name of the principal and the account number, if any, for electronic transfer.
- Include proprietary item certificate if applicable.
- Technical literature to support your product (in technical bid).
- Users' list with contact address in technical bid.

The quote should reach the undersigned on or before 5 pm on or before 24 May, 2017.

The envelope should be marked as **“Refrigerated table top centrifuge”**

To,

Dr. Jonaki Sen
Lab # 11,
Department of Biological Sciences & Bioengineering
Indian Institute of Technology,
Kanpur 208016 (UP)

For any query, contact: Dr. Jonaki Sen

Department of Biological Sciences & Bioengineering
Indian Institute of Technology,
Kanpur 208016 (UP)
Ph: 0512-259-4054

Refrigerated table top centrifuge

Essential components and features:

*Item No. 1: Refrigerated table top centrifuge

*Specification:

1. Table top centrifuge with capacity of 4x250mL.
2. It should have a temp. range -9 to + 40 deg C.
3. It should have a short spin key.
4. It should have a fast cool option & stand by cooling.
5. It should have a facility to put "at set rpm" function.
6. It should have 10 acceleration & 10 breaking ramps.
7. It should have more than 30 program memory.
8. It should have a built in condensation drain to prevent water accumulation & corrosion.
9. It should have the ISO, European CE / UL, IVD certification
10. It should have one fixed angle fully autoclavable rotor with the ability to spin 15 ml and 50 ml falcon tubes at 11,000 rpm.
11. It should have one fixed angle fully autoclavable rotor with the ability to spin 1.5/2 ml tubes at 14,000 rpm.
12. **Warranty period (from date of installation):** Vendors providing Warranty upto 5years will be preferred.