

Indian Institute of Technology Kanpur
Department of Materials Science and Engineering
Enquiry No. KB/IITK/MHRD/RKS-1/AC

Date: Nov. 18, 2011

Closing Date: Nov. 23, 2011

Subject: Supply of Hydrothermal Autoclave

Dear Sir,

We intend to purchase the “Programmable Hydrothermal Autoclave” for our laboratory with detailed specifications given below:

1. Temperature range: Room temperature to 250 ° C (or more).
2. Pressure range: 05-25 bar
3. Time setting is required i.e. programmable autoclave.
4. Warranty required: Two years

Kindly send the quotation(s) with detailed specifications of the item in a sealed envelope to the following address so as to reach to the undersigned on or before the closing date of enquiry.

With best regards,

Dr. Kantesh Balani
Dept. of Materials Science and Engg.
Indian Institute of Technology Kanpur
Kanpur-208016, India