

INDIAN INSTITUTE OF TECHNOLOGY KANPUR

DEPARTMENT OF EARTH SCIENCE

Enquiry No: ES/IS/2014-15/26

Date: 18.02.2015

Subject: Quotation for supply of ANALYTICAL BALANCE :- (BRAND: SARTORIUS, METTLER-TOLEDO).

With reference to the subject mentioned above, you are invited to submit the quotation in a sealed cover in order to reach us by February 25, 2015 in the form of a hard copy to the address mentioned below. If you have any questions please call Dr. I S Sen at 0512-2596440, email: isen@iitk.ac.in.

The prospective suppliers are required to send quotation in two parts in sealed envelopes, as "Technical Bid" and "Financial Bid". The Technical Bid should contain detailed technical specification of the product being offered and should not mention any prices. The Financial Bid should include the detailed price quotation clearly including the cost of the equipment, taxes, service charges if any, shipping and handling charges. **The two separate and sealed envelopes should be clearly marked appropriately as "Technical Bid" and "Financial Bid". Kindly write the inquiry no on the top of envelop.**

Specification for Analytical Balance :- (Brand: Sartorius, Mettler-Toledo)

Maximum Capacity	60-120g
Minimum Capacity	0.01 mg
Repeatability	≤0.05 mg or even better
Settling time	Approx.8s
Connectivity	RS232 to connect to PC, Bluetooth, USB to connect to PC
Calibration	fully automatic time-and temperature- controlled internal adjustment
Size of weighing pan	80-85 mm diameter
Regulatory requirements	ISO or any other relevant trademarks
Operation	keys for main basic functions
Adjustment	Internal, isoCAL

- Overload protection
- Warranty/Guarantee should be clearly mentioned.

Terms and Conditions:

- 1) Maximum education discount, if any should be offered.
- 2) Validity of quotation should be at least for 60 days.
- 3) Prices should be quoted on FOB (Free on Board) only (if imported).
- 4) Normal payment terms for the Institute will be applicable (90% on delivery of the items and the remaining 10% after satisfactory installation/ inspection).
- 5) Quotation should carry proper certifications like agency certificate, proprietary certificate, etc.

Kindly write the inquiry number on the top of envelop.

Indra Sekhar Sen

Dr. Indra Sekhar Sen
WLE 201, Earth Sciences Department
I.I.T. Kanpur PIN 208016