

INDIAN INSTITUTE OF TECHNOLOGY KANPUR

DEPARTMENT OF CHEMISTRY

DATED: 14/10/2015

Kindly send us urgent sealed quotations for the following items against tender number **IITK/CHM/JKB/2015/10**. Quotations should be addressed to Prof. Jitendra K. Bera, Chemistry Department, IIT Kanpur, Kanpur - 208016. Please send technical and financial bids separately. The last date for receipt of sealed quotations is **28/10/2015**.

| <b>S. No.</b> | <b>Item</b>                                       | <b>Specifications</b>   |
|---------------|---|---|
| 1.            | Electronic Analytical Weighing Balance (Imported) | Weighing Capacity: 220 g<br>Readability: 0.1 mg<br>Pan Size: (diameter) 90 mm<br>Response Time: $\leq 2.5$ sec<br>Repeatability: $\leq +0.1$ mg<br>Linearity: $\leq +0.2$ mg<br>Back light Liquid crystal display<br>Automatic Calibration with internal calibration<br>Mass unit conversion options should be available<br>Overload protection should be available<br>Must be an ISO certified product<br>User manual should be included<br>Installation should be done free of cost<br>After sales support should be provided<br>Minimum warranty of one year |