



Dr. Amitabha Bandyopadhyay

Assistant Professor
Convenor, Dept. BSBE
Tele : 91-512-259-4055(Office)
Email: abandopa@iitk.ac.in

Dated: 24-11-2011
Closing date: 30-11-2011

REF – BSBE/AB/NC/2011/05

Request for submission of quotation for Fine Balance

Please send a quotation in triplicate, in a sealed envelope, to the undersigned for the supply of the following:

Fine Balance:

- * Capacity : 220 gms
- * Readability : 0.1 mg
- * Volts : 12 V DC
- * Repeatability Std. Deviation; ≤ 0.1 mg
- * Linearity : ± 0.2 mg.
- * Response Time : 3 secs.
- * Operating ambient Temp. : 5 to 40 Deg. C

Warranty: 3 years

Look forward to your best offer for a model that meets the above

Sincerely,

Dr. Amitabha Bandyopadhyay
Assistant Professor
Dept. of Biological Sciences & Bioengineering
Indian Institute of Technology, Kanpur-208016
INDIA
512-259-4055(Ph)